

INSTRUCTION FOR USE- DOLPHIN STAPLER

DERMAMATE, DERMAMATE ECO AND DURAMATE

SURGICAL DISPOSABLE SKIN STAPLER – 35W

DESCRIPTION:

Dermamate, Dermamate Eco and Duramate Stapler are lever action stapler with excellent staple visibility and is available in a wider staples with 35 staple count.

INTENDED USE

Stapler has the application in surgery for skin closure.

ACTIONS:

Skin staples are made of 316L medical grade stainless steel wire and are non absorbable in nature. It has to be removed after the wound heals through skin staple remover. Cytotoxicity tests shows that there is no tissue toxicity.

APPLICATIONS:

Sutures should be selected and implanted depending on the patient's condition, surgical experience, surgical technique and wound size.

CONTRA – INDICATIONS:

When it is not possible to maintain at least 5 mm distance from the stapled skin to the underlying structures, the use of stapler for skin closure is contra indicated.

The use of this stapler is contraindicated in patients with known sensitivities or allergies to the stainless steel.

It is contraindicated to be used for internal tissue closure i.e. below skin, rectal vaginal, oral cavity and eyes.

WARNINGS/PRECAUTIONS/INTERACTIONS

User should be familiar with surgical procedures and techniques involving non – absorbable stapler and staples before employing stapler and staples for skin closure.

Ensure that the lever (handle) of the stapler is fully compressed for dislodging and forming the staple.

INSTRUCTION FOR USE- DOLPHIN STAPLER

Do not press the stapler's handle twice in a same place or do not press if the staple is not formed.

STERILITY

Sutura[®] sutures are sterilized by ethylene oxide. Do not re-sterilize. Do not use if package is opened or damaged. Discard opened unused sutures.

STORAGE

Recommended storage condition is to store at temperature between 10°C and 35°C, away from moisture and direct heat. Do not use after expiry date.

DISPOSAL:

Discard used stapler contaminated or uncontaminated with blood in the container meant for "infectious waste". Unused expired staplers should be incinerated.



Futura Surgicare Pvt Ltd

86/C2, 3rd Main, 2nd Stage,

Yeshwanthpur Industrial

Suburb, Bangalore, Karnataka,

India, Pin – 560 022

Mfg Lic No: KTK/28/273/1995

E-mail: sales@dolphinsutures.com

Web Site: www.dolphinsutures.com



Obelis s.a

Bd. Général Wahis 53

1030 Brussels, BELGIUM












Tel: + (32) 2. 732.59.54

Fax: +(32) 2.732.60.03

E-Mail : mail@obelis.net

INSTRUCTION FOR USE- DOLPHIN STAPLER

SYMBOLS USED ON THE LABELS:

	Do not re-use		Batch number
	Date of manufacture		Date of expiry
	Sterilized by Ethylene Oxide		Temperature limitation
	Avoid Moisture		Do not re-sterilize
	Do not use if package is damaged or opened		Consult instructions for use
	Avoid direct sunlight		